

FagronLab UVGI-80

Leakage statement





Leakage statement

To whom it may concern,

We hereby confirm that Fagronlab UVGI-80 device is safe for use in the presence of people within working areas for 8 or more hours daily.

The UV leakage is $\leq 5\mu\text{W}/\text{cm}^2$ (measured 30cm from the device) which is not considered to be harmful for people.

It is although recommended that any direct eye contact with the UVC light should be avoided.

Signed by :

Aris Karamperis

Maintenance Engineer FagronLab Development

Fagron Hellas

12km N.R. Trikala – Larisa
P.C. 42100, P.O. Box 32
Trikala, Greece

T +30 24310 83633-5
F +30 24310 83615

www.fagron.gr